

Test Certificate

Part:

1 Liter Vessel inactinic brown

Number:

02503-2


Serial Number:

67390 - 67407

100 mm

200 mm

For SOTAX AT Dissolution Testing Systems



USP/Ph.Eur. Tolerances 98 – 106 mm 160 – 210 mm

With hemispherical bottom

Material:

Dimensions:

Borosilicate glass inactinic brown

а

b

The accuracy of the above mentioned part dimensions have been examined. All relevant measurements are within tolerance and the factory acceptance test has passed. The dimensions are in conformance with USP and Ph. Eur. Pharmacopeia requirements (at time of delivery).

SOTAX Quality Management Bruno Balli 2017-04-07

Inside Ø

Height

Signatui