


Test Certificate

Part: 1 Liter Vessel

Number: 14143-10

Serial Number: 77790 - 77999

For SOTAX AT Dissolution Testing Systems



USP/Ph.Eur. Tolerances

Dimensions:

Inside Ø Height 102.9 to 105.1 mm 200 mm 98 – 106 mm 160 – 210 mm

With hemispherical bottom

b

Material: Borosilicate glass

The accuracy of the above mentioned part dimensions have been examined. All relevant measurements are within tolerance and the factory acceptance test has passed. The dimensions are in conformance with USP and Ph. Eur. Pharmacopeia requirements (at time of delivery).

SOTAX Quality Management Bruno Balli 2019-02-05

CT14143EN-01 1 / 1