

Test Certificate

Part:


1 Liter Vessel

Number:

14143-02

Serial Number:

76809 - 76828



	tenter		USP/Ph.Eur. Tolerances
Dimensions:	Inside Ø Height	a 102.9 to 105.1 mm b 200 mm	98 – 106 mm 160 – 210 mm
	With hemisph	erical bottom	
Material:	Borosilicate g	lass inactinic brown	

The accuracy of the above mentioned part dimensions have been examined. All relevant measurements are within tolerance and the factory acceptance test has passed. The dimensions are in conformance with USP and Ph. Eur. Pharmacopeia requirements (at time of delivery).

SOTAX Quality Management Bruno Balli 2018-11-13