


# Certificate of Compliance

Part Description: **1 Liter Vessel**  
Part Number: 14143-10  
Serial Number: N172001 – N172500

For SOTAX SDT - Dissolution Testing Systems



Parameter	USP <711> Dissolution	Actual
<b>A</b> Inside Ø:	98 – 106	102.90 – 105.10
<b>B</b> Height:	160 – 210	199.30 – 200.70
<b>C</b> Hemispherical bottom:		51.00 – 53.00
Material:		Borosilicate glass

All readings are in mm

The accuracy of the above mentioned part dimensions have been examined in accordance to ISO-Norm ISO 2859. All used instruments are controlled by the SOTAX measuring equipment monitoring according ISO 9001.

All relevant measurements are within tolerance and the factory acceptance test has passed. The dimensions are in conformance with USP Pharmacopeia requirements (at time of delivery).

Inspected

Name: Moreshwar Pulekar

Date: 14-12-2017

Signature:   
14-12-2017 16:07:16